## **Sultan Qaboos University**

**Procurement Department** 



## جامعة السلطان قابوس

دائرة المشتريات

## **REQUISITION FORM**

Requisition Number:7022220046 Last Date for Submission:11/07/2022

S.N.	SQU Product Code	Description	Qty.	UOM	Remarks
1	EQUINSNIT0001	NITRATE METER	1	PCS	The instrument: a handheld device for measuring NO2 (Nitrite) in water with the following specifications: HACH LANGE DR900 OR (OR) Equivalent  Measurement mode: Transmittance (%), Absorbance (Abs) and Concentration (Conc)  Dimensions: (W x D x H) 23.6 x 8.7 x 4.7 cm (9.3 x 3.4 x 1.9 in.)  Enclosure rating: IP67  Weight: 0.6 kg (1.3 Ib)  Power requirements: (internal) AA Alkaline batteries (4x)  Battery life: 6 months (typical) at 5 readings a day/5 day week without the backlight Interface; USB mini  Operating temperature: 0 to 50 ?C (32 to 122 ?F), maximum 90% relative humidity non-condensing  Storage temperature: ?30 to 60 ?C (?22 to 140 ?F), maximum 90% relative humidity non-condensing

Lamp source: Light emitting diode (LED) Detector: Silicon photodiode Wavelength range: 420, 520, 560, 610 nm Photometric measuring range: 0?2 Abs Wavelength accuracy: ?1 nm Photometric accuracy: ?0.005 Abs at 1.0 Abs nominal Photometric linearity: ?0.002 Abs (0?1 Abs) Wavelength selection: Automatic, based on the method selection Photometric repeatability: ?0.002 Abs (0?1 Abs)
the method selection Photometric repeatability: ?0.002 Abs (0?1
AUSI
Data readout: Graphical display, 240 x 160 pixel (backlit)
Sample cells: 16 mm, 1 cm/10 mL, 1-inch (25 mm) round Protection class: Class III
Certifications: CE certified Warranty: 2 years
With following reagents:
Nitrogen, nitrite(No2-N), Nitriver3 quantity 3 (100 measurements)