جامعة السلطان قابوس
Sultan Qaboos University

Requisition Form

Requisition Number: 7038/26/0124

Last Date for Submission: 19/10/2026

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	EQUINSDET0001	DETECTOR	2	Nos	MIC 600 Fixed Gas Detector Main equipment with standard accessories O2: Electrochemical sensor, 0–30% VOL, 0.01% VOL resolution NO2: Electrochemical sensor, 0–500 ppm, 1 ppm resolution NO: Electrochemical sensor, 0–500 ppm, 1 ppb resolution CO2: IR sensor, 0–20000 ppm, 1 ppm resolution CH4: Electrochemical sensor, 0–5000 ppm, 1 ppm resolution Relative RH: 0–95% RH Temperature: -30 ~ 50 Air pressure: 600–1100 hPa warranty 1 year for contact 95705001