



Requisition Form

Requisition Number: 1234/25/0268

Last Date for Submission: 29/12/2025

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	false	Drying Hot Plate	5	Units	<p>For drying microscope slides that have been prepared for staining or other analysis.</p> <p>Temperature Control(0°C-100 °C)</p> <p>Uniform Heating</p> <p>Over-temperature Protection</p> <p>Insulated walls to prevent heat loss</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>
2	EQUHETWAT0001	WATER BATH	2	Nos	<p>For maintaining samples at a consistent temperature by immersing them in heated water</p> <p>Water Tank made of stainless steel</p> <p>Corrosion resistance</p> <p>Temperature Control System (0°C-100°C)with digital controller</p> <p>Sample racks or holders "</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
3	EQUCMDPHR0003	ANALYTICAL BALANCE	4	Nos	<p>Typically measures in increments of 0.001g (1mg) or 0.0001g (0.1mg) for high accuracy.</p> <p>Capacity: 220g</p> <p>Enclosed with a draft shield or weighing chamber</p> <p>Tare Function</p> <p>Auto-Off</p> <p>Touchscreen</p> <p>Readability: 0.1mg</p> <p>Full metal housing with overload protection</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>
4	EQUMEDMIC0001	MICROTOME	4	Nos	<p>"Basic Rotary Microtomes RAZOR "</p> <p>Section thickness range : 0.5–60m</p> <p>Section thickness settings:</p> <p>0.5–2m in 0.5-m steps</p> <p>2–10m in</p> <p>1-m steps</p> <p>10–60m in</p> <p>2-m steps</p> <p>Trim thickness settings : 10m, 20µm, 30µm, 40m</p> <p>Horizontal advance: 28mm</p> <p>Vertical stroke: 60mm</p> <p>Specimen retraction: automatic</p> <p>Specimen orientation: ± 8 °c Dimensions: 560 x 550 x 325mm (width x depth x height) "</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p>

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					<p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>