



Requisition Form

Requisition Number: 1234/25/0269

Last Date for Submission: 29/12/2025

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	false	ID-Card Incubator	2	Units	<p>ID card incubator is a laboratory device used in blood banking for the specific purpose of incubating ID cards (gel cards) at a controlled temperature, typically 37°C. Key specifications include capacity (e.g., 24-30 cards), temperature control (e.g., 40°C $\pm 0.1^{\circ}\text{C}$), a timer for incubation periods (e.g., 15 minutes, adjustable), a digital display for monitoring, an audible alarm to signal completion, and an external design suitable for laboratory benches.</p> <p>Technical specifications Voltage: 100-240V Power: 163VA .Capacity: maximum 24 ID cards .Temperature: 37 °C .Internal dimensions (L) x (W): 20.5 x 20.5 cm .External dimensions (W) x (D) x (H): 33 x 30 x 14 cm</p> <p>- Terms and Conditions:</p> <ol style="list-style-type: none"> 1. Warranty. 2. Delivery period. 3. Installation and commissioning must be done by the supplier. 4. Pre-installation requirements if any.
2	false	Medium Dispenser	1	Units	<p>Basic one : to dispense precise amounts (.5ml -100 ml)of liquid media maintain high accuracy and precision Sterile dispensing systems Volume Control</p> <p>- Terms and Conditions:</p> <ol style="list-style-type: none"> 1. Warranty. 2. Delivery period. 3. Installation and commissioning must be done by the supplier.

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					4. Pre-installation requirements if any.
3	EQUBIOHER0014	INCUBATOR SHAKER	1	Nos	<p>Incubator Shakers: provide precise temperature control, shaking capabilities, and humidity regulation for optimal conditions during experiments</p> <p>Temperature Range: Typically from +5°C above ambient to 70°C</p> <p>Shaking Mechanism: Adjustable speed ranging from 30 to 400 RPM, with an orbital diameter between 15 mm to 50 mm for optimal shaking performance.</p> <p>Over-temperature Protection: Integrated safety features to prevent overheating and ensure sample integrity.</p> <p>Alarm System: Visual and audible alerts for temperature deviations, shaking malfunctions, or other system issues.</p> <p>Versatile Use: Removable shaker functionality, allowing the unit to be used as a flat incubator when shaking is not required</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>
4	false	Micro Pipetting	50	Units	<p>Micro-pipettes:</p> <p>Pipetting and transferring liquids in small, precise amounts 10 µL</p> <p>1. Pipette Volume Range: 0.5 µL to 10 µL</p> <p>- Accuracy: ± 0.1–0.2 µL</p> <p>20 µL</p> <p>2. Pipette Volume Range: 1 µL to 20 µL</p> <p>- Accuracy: ± 0.2–0.3 µL</p> <p>100 µL</p> <p>3. Pipette: Volume Range: 10 µL to 100 µL</p> <p>- Accuracy: ± 0.5–1.0 µL</p>

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					<p>4. 1000 µL Pipette (1 mL):</p> <p>Volume Range: 100 µL to 1000 µL</p> <p>- Accuracy: ± 2.0–5.0 µL</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>
5	false	Hematic Slide Stainer	1	Units	<p>Hematic Slide Stainer: Automated slide staining system for the staining of blood smears, bone marrow samples</p> <p>Slide Capacity, Can process up to 30–40 slides per run.</p> <p>Processing Time of 10–20 minutes for staining cycles.</p> <p>Automatic Reagent Dispensing</p> <p>Automatic cleaning cycle slide racks or holders</p> <p>Digital display: Shows information such as, Staining cycle time, Reagent levels, Slide count.</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>
6	false	Biosafety Cabinet	1	Units	<p>Biosafety Cabinet Class II Type A:</p> <p>Provide a sterile working environment while protecting the user, the laboratory, and the samples from contamination.</p> <p>Airflow and Filtration: Vertical Laminar Flow and exhausted through a HEPA filter.</p> <p>Touchscreen Control Panel</p> <p>Adjustable Front Sash</p>

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					<p>Automatic Self-Testing UV (ultraviolet) for disinfection purposes safety monitoring systems to alert about changes in airflow, filter status, and other critical safety parameters</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>