



Requisition Form

Requisition Number: 1234/25/0270

Last Date for Submission: 29/12/2025

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	false	CBC Analyzer	1	Units	<p>Throughput (samples per hour), parameters reported (e.g., WBC, RBC, HGB, HCT, MCV, PLT, and various differentials), sample volume required, automation level (fully automated with start-up/shut-down features), data storage, integration capabilities (e.g., LIS connectivity), and reagent usage.</p> <p>- Terms and Conditions:</p> <ol style="list-style-type: none">1. Warranty.2. Delivery period.3. Installation and commissioning must be done by the supplier.4. Pre-installation requirements if any.
2	false	Cytospin	1	Units	<p>Cytospin: for cell sedimentation on microscope slides. centrifuge speed ranges from 1,000 to 4,000 rpm hold up to 18 slides Cytospin Chamber adjustable settings for speed, time, and temperature slide holders or slide racks that securely hold the slides in place during centrifugation</p> <p>- Terms and Conditions:</p> <ol style="list-style-type: none">1. Warranty.2. Delivery period.3. Installation and commissioning must be done by the supplier.4. Pre-installation requirements if any.
3	false	Elisa Reader	1	Units	<p>ELISA Reader: plate reader to measure the optical density (OD) of samples in ELISA tests Wavelength Range: (from around 400 nm to 800 nm)</p>

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					<p>Microplate Capacity 96-well plates Detection Mode, Fluorescence Mode, Luminescence Mode offer high sensitivity and precision</p> <p>- Terms and Conditions:</p> <p>1. Warranty.</p> <p>2. Delivery period.</p> <p>3. Installation and commissioning must be done by the supplier.</p> <p>4. Pre-installation requirements if any.</p>