



جامعة السلطان قابوس  
Sultan Qaboos University

## Requisition Form

Requisition Number: 1234/26/0090

Last Date for Submission: 27/07/2026

| S.N. | SQU Product Code | Description                   | Qty | UOM   | Remarks                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|------|------------------|-------------------------------|-----|-------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1    | SPRSPE0178       | biosafety Cabinet Maintenance | 7   | Units | includes: Supply and Replacement of HEPA Filter - Supply and Replacement of Pre-filter - On-site Complete validation in accordance with ISO 14644 or EN 12469 manufacture specification, CETA, USP 797, GMP Standard manufacture specification including the following tests: * Ceiling HEPA/ULPA Filter Integrity test. * Airborne Particle Counting Test. * Air velocity test. * Airflow Visualization Studies (Smoke test). * Temperature and Humidity Test. * Light intensity test. * Noise level test. * Validation including report with EIAC ISO 17020 Accredited Certificate. |