



جامعة السلطان قابوس
Sultan Qaboos University

Requisition Form

Requisition Number: 7022/26/0011

Last Date for Submission: 11/05/2026

| S.N. | SQU Product Code | Description | Qty | UOM | Remarks |
|------|------------------|-------------------|-----|----------------|---|
| 1 | EQULABPLA | SURGICAL PLATFORM | 1 | Not Applicable | <p>Electrosurgical Unit (ESU) High-frequency electrosurgical generator for soft tissue cutting and coagulation in veterinary surgery. For cutting and coagulating tissue during surgery, reducing bleeding. Output Power: 300–400 W (adjustable for monopolar and bipolar modes) Modes: Monopolar cut Blend / coagulation Bipolar cut and coagulation Spray coagulation mode Frequency Range: 400–500 kHz Control Interface: Digital touchscreen or membrane keypad with LED indicators Safety Features: Automatic power adjustment based on tissue impedance Audible and visual alarms for errors or high load Return electrode monitoring system (REM) Overheat and overcurrent protection Accessories & Instruments Included: Monopolar handpiece with various tips Bipolar forceps and cable Footswitch (dual pedal for cut and coag) Patient plate (reusable or disposable) Power cable and electrode holder Mobility: Tabletop or tower-mount compatible Power Supply: 220–240 V / 50–60 Hz Certifications: CE / ISO 13485 / Veterinary use approved Accessories: (Basic Surgical Instrument Set (4 sets)+ Orthopedic Instrument Set (small animals) (2 sets)+ Dental Surgery Instrument Set including scalers (small animals) (1 set)+ Dental Instrument Set (Equine) with power float (1set) with mouth gag and mouth flush manual system+ Ophthalmic Instrument Set (1set)</p> |

| S.N. | SQU Product Code | Description | Qty | UOM | Remarks |
|------|------------------|----------------------------|-----|--------|--|
| | | | | | <p>+ Teat Surgery Instrument Set (1 set)+ Dental Unit Small Animals (High speed) (1 PC) Voltage: 220V+-20V/50Hz - British Standard Electrical Plug Terms and Conditions: The quotation must include the following details: 1. Warranty period. 2. Delivery period. 3. Price List for the major spare parts. 4. Cost of maintenance visit per day after the warranty period expires. 5. Submission a copy of an Operation; Service manuals at the installation time. 6. Back up/spare parts support up to the 10th year. 7. Details of any pre-installation requirements (attached with offer). 8. Manufacturing date certificate from the manufacturer. 9. Delivery, Installation, commissioning; on-site training to be done by the supplier.</p> <p>*Note: - Bid evaluation criteria will be subjected to; Price, specifications, warranty period, fast delivery, known brand; after-sales technical support. - Since this is a non-profit organization and the item to be used for educational purposes, a maximum discount should be given as far as possible.</p> |
| 2 | EQUINSAST | MOBILE ANAESTHETIC MACHINE | 2 | Pieces | <p>Veterinary Anesthesia Machine with Vaporizer (Small animals) Brand: Mindray - Veta 5Plus or Equivalent. Breathing Synchronization, Automatic adjustment of mechanical ventilation parameters, Effective prevention of hypoventilation with automatic backup ventilation. With monitor. Veterinary Anesthesia Machine with Oxygen Concentrator For safe inhalation anesthesia in dogs, cats, and small animals. Includes vaporizer (Isoflurane or Sevoflurane, agent-specific, temperature-compensated) Dual flowmeter (O 0–4 L/min, optional NO) Rebreathing and non-rebreathing circuits (Bain & circle systems) Integrated mechanical ventilator (manual/auto modes, adjustable tidal volume & rate) O flush valve, pressure gauge, APL valve, and CO absorber Mounted on mobile stand with</p> |

| S.N. | SQU Product Code | Description | Qty | UOM | Remarks |
|------|------------------|-------------|-----|-------|---|
| | | | | | <p>wheels and brakes Accessories included: Oxygen Concentrator Endotracheal tubes (sizes 2.0 – 10.0 mm ID) Breathing bags (0.5 L, 1 L, 2 L, 3 L) Breathing circuits (rebreathing & Bain) Scavenging interface and O hose Reservoir bag mount and connectors Power supply: 220–240 V / 50–60 Hz CE/ISO certified veterinary-grade system Voltage: 220V+-20V/50Hz - British Standard Electrical Plug Terms and Conditions: The quotation must include the following details: 1. Warranty period. 2. Delivery period. 3. Price List for the major spare parts. 4. Cost of maintenance visit per day after the warranty period expires. 5. Submission a copy of an Operation; Service manuals at the installation time. 6. Back up/spare parts support up to the 10th year. 7. Details of any pre-installation requirements (attached with offer). 8. Manufacturing date certificate from the manufacturer. 9. Delivery, Installation, commissioning; on-site training to be done by the supplier.</p> <p>*Note: - Bid evaluation criteria will be subjected to; Price, specifications, warranty period, fast delivery, known brand; after-sales technical support. - Since this is a non-profit organization and the item to be used for educational purposes, a maximum discount should be given as far as possible.</p> |
| 3 | false | Endoscope | 1 | Units | <p>Portable veterinary endoscope for large animals (rigid and flexible). 1.5 meter -12 mm. with ports for pyopsy channel (water and gas) Brand: Medequus or Olympus or Equivalent Portable endoscopy+ Gastroscope+ Dental Scope bundle 1)GASTRO video (Gastroscope) - 11mm outer diameter- 330cm working length- 2.8mm working channel- 120 degree FOV- Integrated high power LED light source</p> |

| S.N. | SQU Product Code | Description | Qty | UOM | Remarks |
|------|------------------|-------------|-----|-----|--|
| | | | | | <p>- 7 inch touch screen, imaging hub Microchip reader enabled Additional accessories:- Gastroflator Pump- Battery Powered Suction Pump- Roller Case Smart Clamp Scope Holder- Scope Magic Arm- Cut off Stomach Tube, 1.2m- Plus included cleaning accessories</p> <p>2) Video Endoscope (Flexible) 9mm outer diameter, 160cm working length, 2.8mm working channel, portable endoscope with flush- 2.8mm working channel- Lens flushing channel- Portable endoscope- Integrated high power LED light source- Laser compatible Additional accessories:- Guttural Pouch Probe- Included cleaning accessories</p> <p>3) Rigid dental scope specifications include: - 10mm diameter - 70 degree angle of view - 45 cm working length - 320 degree rotatable - Freeze/capture button - Integrated LED light source - Auto focus - Luer flushing channel Additional accessories: 1. Biopsy forceps (2.3–2.8 mm) – Routine gastrointestinal mucosal biopsy in horses, dogs, and cats. 2. Alligator grasping forceps (2.8 mm) – Foreign body retrieval from esophagus and stomach. 3. Polypectomy snare (2.3 mm) – Removal of polyps and abnormal soft tissue. 4. Cytology brush (2.0–2.3 mm) – Collection of mucosal samples for cytology. 5. Suction & irrigation tubing set (standard) – Removal of blood, mucus, and debris during endoscopy. 6. Water bottle with tubing (1–2 L) – Lens cleaning and flushing during procedures. 7. Leak tester (flexible endoscope type) – Integrity testing before cleaning and disinfection. 8. High-level disinfection tray/system – Proper cleaning and disinfection of flexible endoscopes. 9. Endoscope storage cabinet (vertical hanging) – Safe storage and prevention of scope damage. 10. Protective transport case – Safe movement of endoscope between locations. 11. Disposable bite blocks (small & large sizes) – Protection of endoscope during oral procedures.</p> |

| S.N. | SQU Product Code | Description | Qty | UOM | Remarks |
|------|------------------|-------------|-----|-----|---|
| | | | | | <p>12. Dental Scope Holder. Plus included cleaning accessories Voltage: 220V+-20V/50Hz - British Standard Electrical Plug Terms and Conditions: The quotation must include the following details: 1. Warranty period. 2. Delivery period. 3. Price List for the major spare parts. 4. Cost of maintenance visit per day after the warranty period expires. 5. Submission a copy of an Operation; Service manuals at the installation time. 6. Back up/spare parts support up to the 10th year. 7. Details of any pre-installation requirements (attached with offer). 8. Manufacturing date certificate from the manufacturer. 9. Delivery, Installation, commissioning; on-site training to be done by the supplier.</p> <p>*Note: - Bid evaluation criteria will be subjected to; Price, specifications, warranty period, fast delivery, known brand; after-sales technical support. - Since this is a non-profit organization and the item to be used for educational purposes, a maximum discount should be given as far as possible.</p> |