



جامعة السلطان قابوس
Sultan Qaboos University

Requisition Form

Requisition Number: 7022/26/0040

Last Date for Submission: 04/05/2026

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	EQUANAMIC0001	MICROSCOPE	4	Nos	<p>[EQUANAMIC0001] MICROSCOPE MICROSCOP Technical Specifications: – Optical System: Infinity optical system – Observation Tube: 30° inclined binocular tube Interpupillary distance adjustment range: 48 – 75 mm, Eyepoint adjustment: 370.0 – 432.9 mm – Eyepiece (10x): Field Number (FN): 20 (anti-fungus) – Revolving Nosepiece: Fixed quadruple nosepiece with inward tilt – Objectives: Infinity Plan achromat, anti-fungus 4x NA: 0.10 W.D.: 27.8 mm 10x NA: 0.25 W.D.: 8.0 mm 40x NA: 0.65 W.D.: 0.6 mm 100x Oil NA: 1.25 W.D.: 0.13 mm – Focusing: Stage height movement (coarse movement stroke: 15 mm), coarse adjustment limit stopper, Torque adjustment for coarse adjustment knob, Fine focus knob (minimum adjustment gradations: 2.5 µm) – Stage: Wire movement mechanical xed stage Traveling range: 76 mm (X) x 30 mm (Y), Specimen holder, Specimen position scale – Condenser: Abbe condenser NA 1.25 with oil immersion – Illumination System: Built-in transmitted illumination system, LED Power Consumption 0.5 W (nominal values) – Body: Aluminum die-casting metal frame, Protective covering.</p> <p>The quotation must include the following details:</p> <ol style="list-style-type: none">1. Warranty period.2. Delivery period.3. Price List for the major spare parts.4. Cost of maintenance visit per day after the warranty period expires.5. Submission a copy of an Operation; Service manuals at the installation time.6. Back up/spare parts support up to the 10th year.7. Details of any pre-installation requirements (attached with offer).8. Manufacturing date certificate

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
					from the manufacturer. 9. Delivery, Installation, commissioning; on-site training to be done by the supplier. -