



جامعة السلطان قابوس
Sultan Qaboos University

Requisition Form

Requisition Number: 7038/26/0099

Last Date for Submission: 17/08/2026

S.N.	SQU Product Code	Description	Qty	UOM	Remarks
1	false	Deionized Water System for HPLC	1	Units	Water purification system for the production of Type I (ultrapure) water for HPLC, Resistivity: 18.2 M·cm at 25°C, Total Organic Carbon (TOC): 5 ppb.